

LOM07 Hybrid 2021

Hybrid Baseline Configuration Tests




12 May 2021

LOM07 Hybrid baseline_config Test:

Power supply readings after loading baseline_config_L0.xml:

4/4 APVs Synced
(V) | (A)
AVDD: 2.50 | 0.327
DVDD: 2.50 | 0.255
V125: 1.25 | 0.234
FPGA: 5.52 | 1.15

Baseline config run successful.



LOM07 hybrid icalScal:

Successful at 120Hz scan

Sample 3 Data

