

LOM08 Hybrid 2021

Hybrid Baseline Configuration Tests




12 May 2021

LOM08 Hybrid baseline_config Test:

Power supply readings after loading baseline_config_L0.xml:

4/4 APVs Synced
(V) | (A)
AVDD: 2.50 | 0.287
DVDD: 2.50 | 0.254
V125: 1.25 | 0.234
FPGA: 5.52 | 1.16

Baseline config run successful.



LOM08 hybrid icalScal:

Successful at 1000Hz scan

Sample 3 Data

