

L0M11 Hybrid 2021

Hybrid Baseline Configuration Tests




14 May 2021

L0M11 Hybrid baseline_config Test:

Power supply readings after loading baseline_config_L0.xml

4/4 APVs Synced
(V) | (A)
AVDD: 2.50 | 0.344
DVDD: 2.50 | 0.255
V125: 1.25 | 0.238
FPGA: 5.52 | 1.14

Baseline config run successful.



LOM11 hybrid icalScal:

Successful at 120Hz scan

Sample 3 Data

