

L0M24 Hybrid 2021


Hybrid Baseline Configuration Tests

27 May 2021


L0M24 Hybrid baseline_config Test:
Power supply readings after loading baseline_config_L0.xml
4/4 APVs Synced
(V) | (A)
AVDD: 2.50 | 0.292
DVDD: 2.50 | 0.252
V125: 1.25 | 0.238
FPGA: 5.52 | 1.14

Baseline config run successful.


Sample 0 Data



Sample 3 Data



Sample 3 Noise



LOM24 hybrid icalScal:

Successful at 120Hz scan

Sample 3 Data

